Document Reference: Document Issue Number: Document Issue Date: Page 1 of 2 ME-EUDOC-018 06 27/07/2022



## **EU Declaration of Conformity**

This European Declaration of Conformity is issued under the sole responsibility of the manufacturer.

MANUFACTURER				
Name of Company	Registered Office	Trading Address	SRN	
Medicare Products Ltd	South Fen Road, Bourne, Lincolnshire, PE10 ODN. UK	Unit B, Dolphin Way, Purfleet, Essex, RM19 1NZ, UK	GB-MF-000011612	

AUTHORIZED REPRESENTATIVE				
Name of Company	Address	SRN	Phone/email	
International Associates Auditing & Certification Limited	The Black Church, St Mary's Place, Dublin 7, D07 P4AX Ireland	IE-AR-000002248	<u>+353 16971561</u> EUAR@ie.ia-net.com	

Product Name	Code / Catalog Number	Basic UDI-DI
Indigo Nitrile Powder Free Examination Glove (Nitrex Accelerator Free)	GN06	506004079GNPFNS00LX
Intended Purpose	Photo	
Examination gloves for non-invasive use to prevent contact with skin, bodily fluids, and chemicals. Invasive with respect to natural body orifices.		

RISK CLASS FOR MEDICAL DEVICES			
<b>Device Classifie</b>	ation	Common Specifications / Standards	
Class:	I	EN455-1:2020 EN455-2:2015	Medical gloves for single use: Freedom from holes Medical gloves for single use: Physical properties
Rule:	5	EN455-3:2015 EN455-4:2009	Medical gloves for single use: Biological evaluation Medical gloves for single use: Shelf-life determination

## **RISK CATEGORY OF PERSONAL PROTECTIVE EQUIPMENT AND PERFORMANCE LEVELS** Product Risk Category Personal Protective Equipment (PPE) Category III Standard **Performance Levels** EN ISO 374-1:2016 Туре В – К, Т, Р Protection against bacteria and fungi Protection against viruses EN ISO 374-5:2016 Pass Pass EN388:2016 + A1:2018 N/A EN407:2020 N/A EN511:2006 N/A

T: +44 (0) 1708 863868 E: enquiries@medicareproducts.com W. www.medicareproducts.com

Document Reference: Document Issue Number: Document Issue Date: Page 2 of 2 ME-EUDOC-018 06 27/07/2022



Medicare Products Ltd declares that the above-mentioned product meets the provision of the following EU legislation:

- Medical Devices Regulation (EU) 2017/745
- Is classed as Cat III PPE in conformity with the provisions of Regulation (EU) 2016/425 on personal protective equipment and with National Standards transposing harmonised standards EN 374-1:2016+A1:2018, EN 374-5:2016, and EN ISO 21420:2020.
- Is identical to the personal protective equipment which is the subject of EU Type-Examination Certificate number 2777/10013-05/E03-01, issued by SATRA Technology Europe Limited, Bracetown Business Park, Clonee, Dublin, D15 YN2P, Republic of Ireland (Notified Body number 2777), according to Annex V (Module B) of Regulation (EU) 2016/425.
- Is subject to the quality assurance conformity of the production process Module D set out in Annex VIII of Regulation (EU) 2016/425, under the surveillance of the notified body BSI Group The Netherlands B.V., Say Building, John M. Keynesplein 9, 1066 EP, Amsterdam (Notified Body Number 2797).

COMPANY REPRESENTATIVE:	David Langridge		
TITLE:	Head of Technical	SIGNATURE:	D Longridge
PLACE:	Bourne, UK	ISSUE DATE:	27 July 2023
EXPIRY DATE:	27 July 2028		

T: +44 (0) 1708 863868 E: enquiries@medicareproducts.com W. www.medicareproducts.com